IVD solutions through partnership



CHROMagar[™] E.coli

For detection and enumeration of *E.coli*



• CHROMagar[™] E.coli




Plate Reading

- E. coli
- \rightarrow blue
- Other gram negative bacteria
- \rightarrow colourless
- Gram positive
- \rightarrow inhibited



Manufacturer:

CHROMagar 4 place du 18 juin 1940 75006 Paris - France e-mail: CHROMagar@CHROMagar.com www.CHROMagar.com

Distributed by: Mast Diagnostica GmbH Feldstraße 20 DE-23858 Reinfeld Tel: +49 (0)4533 2007 0 Fax: +49 (0)4533 2007 68 e-mail: mast@mast-diagnostica.de www.mast-group.com

For detection and enumeration of *E.coli* in food, water and environmental samples

Background

Contamination by faecal material from animals can be shown by the detection of Escherichia coli in the sample. *E. coli* can contaminate drinking water when the water treatment system is inadequate or during periods of very high rainfalls.

Monitoring of food and water production is essential. High contamination may lead to the suspension of the water supply and food recall by supermarkets.

- Concerning bathing water, regulations are more and more strict:
 - European directive from 1976: 2.000 Escherichia coli (E.coli) bacteria for 100 mL of water.
 - New directive in 2006: 500 E.coli per 100 mL.

The presence of *E. coli* indicates faecal contamination and potential presence of dangerous pathogens such as bacteria like *Vibrio cholerae, Salmonella, Pseudomonas* etc., or viruses and intestinal parasites. The infections resulting from ingestion of contaminated matter can be dangerous and life-threatening.

Medium Performance

18-24h detection

Easy reading and interpretation

The general food and water standards limits' are usually from zero to single figure *E. coli* CFU per gram and thus it is important to detect and enumerate them accurately. With CHROMagarTM E. coli, colonies of *E. coli* develop with an intense blue colour - thus making detection and enumeration of this important hygiene indicator as simple as possible.

2 Lighter workload

Traditional *E. coli* detection methods are extremely tedious and labor-intensive, requiring studies of many colonies.

Qality

(3)

CHROMagar™E.coli media contain 5 % more agar than other media on the market. This helps considerably with the application and streaking of the sample onto the plate.

The media is also suitable for the membrane filtration technique or the pouring technique.

Medium Description

Powder Base	Total
Usual Samples	Processed food, raw materials, water, milk & environment
Procedure	Pouring, Isolation or membrane filtration technique. Incubation 24 h, 37 °C. Aerobic conditions.
Scientific Publications on this product: available on www.CHROMagar.com Please read carefully the instructions for use (IFU document) available on www.CHROMagar.com	

Ordering Information

5	
Product	Order Code
CHROMagar™ E.Coli dry media, 5 liter	15EC168
CHROMagar [™] E. Coli ready to use plates, 20 pcs.	201418

Version 5.0 /